Document reference no.: 152302/6

Report no.: RN-0054



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

CD97

TDS Economical Pocket Tester

Are in compliance with:

Directive 2011/65/EU

RoHS2

Directive 2014/30/EU

EMC and its standards

Directive 2014/35/EU

Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or references to the other technical specification:

EN 50581:2012

Restriction of the use of certain hazardous substances (RoHS)

EC Regulation 1907/2006 EN 61326-1: 2013 Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH)

Electrical Equipment of Measurement, Control and Laboratory use - EMC requirements Part 1: General requirements and its standards:

EMC requirements Part 1:

General requirements and its standards:

EN 61000-4-2: 2009, EMC

Electrostatic Discharge Immunity Test

4 kV/8 kV contact / air

Class A

EN 61000-4-3: 2006, EMC

Radiated RF, Electromagnetic Field Immunity Test

3 V/m (80 MHz to 1 GHz)

EN 61010-1: 2010 (IEC 61010-1: 2010)

Safety Requirements for Electrical Equipment for Measument, Control and

Laboratory Use Part 1: General Reguirements.

 $The \ manufacturer is \ not \ liable \ for \ modifications \ made \ by \ the \ user \ which \ may \ alter \ the \ device's \ conformity \ with \ the \ CE \ label, \ nor \ the \ consequences \ thereof.$

This declaration is the responsibility of the manufacturer:

Milwaukee Electronics Kft.

Alsó-kikötő sor 11/C

H-6726 Szeged - HUNGARY

tel: +36 62 428050

fax: +36 62 428 051

www.milwaukeeinst.com

Milwaukee Instruments, Inc.

(United States of America only)

2950 Business Park Drive - Rocky

Mount NC 27804 - U.S.A

tel: +1 252 443 3630

fax: +1 252 4431937 www.milwaukeeinstruments.com CE

Test Laboratory

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

Milwaukee Instruments, Inc.

Test performed by:

Kovacs Lorand / ETO Eng. /

[Name / Title of Signatory / Signatu

Declaration authorized by: Vasile Prodan / ETO Responsible /

[Name/Title of Signatory/Signature]

Issue place / Date:

Nuşfalău / May 23, 2018

[Place / Date in format: Month, dd, yyyy]

FO-MWCETEST r.3 - 2018.05.14

Page 1 of 1