Document reference no.: 153816/1 Report no.: RN-0090



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:MI106pH / Temperature Professional Portable Meter

Are in compliance with:

Directive 2011/65/EU Directive 2014/30/EU Directive 2014/35/EU RoHS2 EMC and its standards Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or references to the other technical specification: EN 50581:2012 Restriction of the use of certain hazardous substances (RoHS) EC Regulation 1907/2006 Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH) EN 61326-1: 2013 Electrical Equipment of Measurement, Control and Laboratory use Class A - EMC requirements Part 1: General requirements and its standards: EMC requirements Part 1: General requirements and its standards: EN 61000-4-2: 2009, EMC Electrostatic Discharge Immunity Test 4 kV/8 kV contact / air EN 61000-4-3: 2006, EMC Radiated RF, Electromagnetic Field Immunity Test 10 V/m (80 MHz to 1 GHz) EN 61010-1: 2010 (IEC 61010-1: 2010) Safety Requirements for Electrical Equipment for Measument, Control and Laboratory Use Part 1: General Reguirements.

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

| Milwaukee Electronics Kft. Alsó-kikötő sor 11/C H-6726 Szeged - HUNGARY tel: +36 62 428050 fax: +36 62 428 051 | Milwaukee Instruments, Inc. (United States of America only) 2950 Business Park Drive – Rocky Mount NC 27804 - U.S.A tel: +1 252 443 3630 |
|---|--|
| www.milwaukeeinst.com | fax: +1 252 4431937 |
| | www.milwaukeeinstruments.com |
| Test Laboratory | Milwaukee Instruments, Inc. |
| This is the result of test that was carried out from the submitted type-samples of a product in | Test performed by: Kovacs Lorand / ETO Eng. / [Name / Title of Signatory / Signature / / / |
| conformity with the specification of the respective standards. | Declaration authorized by: Vasile Prodan / ETO Responsible / |
| The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample. | [Name / Title of Signatory / Signature] Issue place / Date: Nuşfalău / May 23, 2018 [Place / Date in format: Month, dd, yyyy] |

FO-MWCETEST r.3 - 2018.05.14

Page 1 of 1