

EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

MI160	pH / ORP / ISE / Temperature Laboratory Bench Meter, 230V
MI160-US	pH / ORP / ISE / Temperature Laboratory Bench Meter, 115V

Are in compliance with:	Directive 2011/65/EU	RoHS2
	Directive 2014/30/EU	EMC and its standards
	Directive 2014/35/EU	Low voltage and its standards

The instruments are in conformity with the following harmonized standards used or references to the other technical specification:

EN 50581:2012	Restriction of the use of certain hazardous substances (RoHS)	
EC Regulation 1907/2006	Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH)	
EN 61326-1: 2013	Electrical Equipment of Measurement, Control and Laboratory use - EMC requirements Part 1: General requirements and its standards:	Class A
EMC requirements Part 1:	General requirements and its standards:	
EN 61000-4-2: 2009, EMC	Electrostatic Discharge Immunity Test	4 kV/ 8 kV contact / air
EN 61000-4-3: 2006, EMC	Radiated RF, Electromagnetic Field Immunity Test	3V/ m (80 MHz to 1 GHz)
EN 61010-1: 2010 (IEC 61010-1: 2010)	Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use Part 1: General Requirements.	
EN 61000-4-4: 2012, EMC	Electrical Fast Transient/Burst Immunity Test	1 kV (5/50 ns, 5 kHz)

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

Milwaukee Electronics Kft.

Alsó-kikötő sor 11/C
H-6726 Szeged - HUNGARY
tel: +36 62 428050
fax: +36 62 428 051
www.milwaukeeinst.com

Milwaukee Instruments, Inc.

(United States of America only)
2950 Business Park Drive - Rocky
Mount NC 27804 - U.S.A
tel: +1 252 443 3630
fax: +1 252 4431937
www.milwaukeeinstruments.com



Test Laboratory

This is the result of test that was carried out from the submitted type-samples of a product in conformity with the specification of the respective standards.

The declaration holder has the right to fix the CE-mark for LVD on the product complying with the inspection sample.

Milwaukee Instruments, Inc.

Test performed by: Kovacs Lorand / ETO Eng. /
[Name / Title of Signatory / Signature]

Declaration authorized by: Vasile Prodan / ETO Responsible /
[Name / Title of Signatory / Signature]

Issue place / Date: Nuşfalău / May 23, 2018
[Place / Date in format: Month, dd, yyyy]