Document reference no.: 171110/1 Report no.: RN-0112



## **EU DECLARATION OF CONFORMITY**

## This certifies that the following designated products:

MI490

Photometer for the determination of PEROXIDE VALUE in the process of oil making

Are in compliance with:

Directive 2011/65/EU Directive 2014/30/EU Directive 2014/35/EU RoHS2 EMC and its standards Low voltage and its standards

|  | The instruments are in conformity with the following harmonized standards used or references to the other technical specification: |  |                          |
|--|--|--|--------------------------|
|  | EN 50581:2012  | Restriction of the use of certain hazardous substances (RoHS)                        |                          |
|  | EC Regulation 1907/2006  | Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH)         |                          |
|  | EN 61326-1:2013  | Electrical Equipment of Measurement, Control and Laboratory use                      | Class A                  |
|  |  | <ul> <li>EMC requirements Part 1: General requirements and its standards:</li> </ul> |                          |
|  | EMC requirements Part 1:   | General requirements and its standards:  |                          |
|  | EN 61000-4-2: 2009, EMC  | Electrostatic Discharge Immunity Test  | 4 kV/ 8 kV contact / air |
|  | EN 61000-4-3: 2006, EMC  | Radiated RF, Electromagnetic Field Immunity Test                                     | 6 V/ m (80 MHz to 1 GHz) |
|  | EN 61010-1: 2010 (IEC 61010-1: 2010)   | Safety Requirements for Electrical Equipment for Measument, Control and              |                          |
|  |  | Laboratory Use Part 1: General Reguirements.   |                          |

The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

## This declaration is the responsibility of the manufacturer:

| <b>Milwaukee Electronics Kft.</b><br>Alsó-kikötő sor 11/C<br>H-6726 Szeged - HUNGARY<br>tel: +36 62 428050<br>fax: +36 62 428 051<br>www.milwaukeeinst.com | Milwaukee Instruments, Inc.<br>(United States of America only)<br>2950 Business Park Drive – Rocky<br>Mount NC 27804 - U.S.A<br>tel: +1 252 443 3630<br>fax: +1 252 4431937<br>www.milwaukeeinstruments.com |
|--|---|
| Test Laboratory  | Milwaukee Instruments, Inc.   |
| This is the result of test that was carried out<br>from the submitted type-samples of a product in   | Test performed by: Kovacs Lorand / ETO Eng. /<br>[Name / Title of Signature]  |
| conformity with the specification of the respective standards.   | Declaration authorized by: Vasile Prodan / ETO Responsible / / / / / / / / / / / / / / / / / / /  |
| The declaration holder has the right to fix the<br>CE-mark for LVD on the product complying with<br>the inspection sample.                                 | Issue place / Date: Nuşfalău / May 23, 2018<br>[Place / Date in format: Month, dd, yyyy]  |

FO-MWCETEST r.3 - 2018.05.14