Milwaukee Electronics Kft. Alsokikoto sor 11. H-6726 Szeged - HUNGARY Tel.: +36 62 428 050; Fax: +36 62 428 051 Milwaukee Instruments, Inc. 2950 Business Park Drive Rocky Mount, NC 27804 – U.S.A. Tel.: +1.252.443.3630; Fax: +1.252.443.1937

## Milwaukee

Document No: RN-0028

## **CE DECLARATION OF CONFORMITY**

*This certifies that the following designated products* 

## **Mi605**

are in conformity with the EMC directive 2004/108/EC and its standards, and Low Voltage Directive 2006/95/EC and its standards:

EN 61326-1: 2006 Electrical Equipment for Measurement, Control and Laboratory use
EMC requirements Part 1: General requirements and its standards:
EN 61000-4-2: 2009, EMC - Electrostatic Discharge Immunity Test

EN 61000-4-3: 2006, EMC - Radiated RF, Electromagnetic Field Immunity Test EN 61000-4-4: 2004, EMC - Electrical Fast Transient/Burst Immunity Test

**EN 61010-1: 2001 (IEC 61010-1: 2001)** Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use Part 1: General Requirements.

Date of issue:

June 15, 2012

Name and Function:

Marius Albert elec.eng. N. Moring

Class A

4 kV/8 kV contact/air 3 V/m (80 MHz to 1 GHz) 0.5 kV (5/50 ns, 5 kHz)

Rev. 0.1 2012-06-15