Document reference no.: 152302/3 Report no.: RN-0051



EU DECLARATION OF CONFORMITY

This certifies that the following designated products:

TH300

Thermistor Thermometer with Stainless Steel Probe and 1 meter cable

Are in compliance with:

Directive 2011/65/EU Directive 2014/30/EU Directive 2014/35/EU RoHS2 EMC and its standards Low voltage and its standards

The instruments are in conformity with	the following harmonized standards used or references to the other technical sp	ecification:
EN 50581:2012	Restriction of the use of certain hazardous substances (RoHS)	
EC Regulation 1907/2006	Registration, Evaluation, Authorisation and Restriction of Chemicals (REACH)	
EN 61326-1: 2013	Electrical Equipment of Measurement, Control and Laboratory use	Class A
	 EMC requirements Part 1: General requirements and its standards: 	
EMC requirements Part 1:	General requirements and its standards:	
EN 61000-4-2: 2009, EMC	Electrostatic Discharge Immunity Test	4 kV/ 8 kV contact / air
EN 61000-4-3: 2006, EMC	Radiated RF, Electromagnetic Field Immunity Test	10 V/ m (80 MHz to 1 GHz)
EN 61010-1: 2010 (IEC 61010-1: 2010)	Safety Requirements for Electrical Equipment for Measument, Control and	
	Laboratory Use Part 1: General Reguirements.	

* The manufacturer is not liable for modifications made by the user which may alter the device's conformity with the CE label, nor the consequences thereof.

This declaration is the responsibility of the manufacturer:

i his declaration is the	responsibility of the manufacturer:	
Milwaukee Electronics Kft.	Milwaukee Instruments, Inc.	
Alsó-kikötő sor 11/C	(United States of America only)	
H-6726 Szeged - HUNGARY	2950 Business Park Drive – Rocky	
tel: +36 62 428050	Mount NC 27804 - U.S.A	
fax: +36 62 428 051	tel: +1 252 443 3630	
www.milwaukeeinst.com	fax: +1 252 4431937	
	www.milwaukeeinstruments.com	
Test Laboratory	Milwaukee Instruments, Inc.	
This is the result of test that was carried out from the submitted type-samples of a product in	Test performed by: Kovacs Lorand / ETO Eng. / [Name / Title of Signatory / Signature]	
conformity with the specification of the respective standards.	Declaration authorized by: Vasile Prodan / ETO Responsible /	
The declaration holder has the right to fix the	[Name / Title of Signatory / Signature]	
CE-mark for LVD on the product complying with	lssue place / Date: Nuşfalău / May 23, 2018	
the inspection sample.	[Place / Date in format: Month, dd, yyyy]	

FO-MWCETEST r.3 - 2018.05.14

Page 1 of 1